

# ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

System: Skyrad NIP

Site: Nauru

## OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 450)X10-6	Offset	Date of change UTC	Time Start UTC	Time End UTC
NIP	31349E6	Skyrad	8 July 2004	8.1091		26 Sept 2004	00:30	01:35

## NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 450)X10-6	Offset	Copy Config. Logger to PC
NIP	31350E6	Skyrad	26 Sept 2004	8.3338		yes

Replacement completed by: Troy Culkan & Rex Pearson

## NOTES:

Old Part remains on site as spare: ☐

Old part packed for shipment to: SGP

Other old part disposition (explain):

Other NOTES:  
FM(OPS)-008.001